

CIRAS-3SCDC Console Software

The CIRAS-3SCDC Console Software is provided as a file named "Ciras3SCDC_Console_vxxx.cab". It is easiest to follow these instructions via the use of a USB mouse attached to the console.

Important Note. If the CIRAS3SCDC Console Software (Ciras3_Console.exe) is already installed, you must uninstall it prior to loading new versions.

Uninstall Instructions

1. Open "My Device" on the console desktop. Navigate to folder: "\Flash Disk\InstalledFiles\PP Systems CIRAS3SCDC_Console_vxxx" Where "vxxx" is the prior installed version.
2. Copy "PP Systems CIRAS3SCDC_Console_vxxx.unload" file.
3. Navigate to "\Windows" folder and paste the unload file.
4. Close "My Device" window.
5. Go to Start>Setting>Control Panel>Remove Programs. Select "PP Systems CIRAS3SCDC_Console_vxxx" from the list and click Remove.
6. After uninstall is complete, close Control Panel. Use "My Device" and navigate to: \Flash Disk\InstalledFiles". Delete folder "PP Systems CIRAS3SCDC_Console_vxxx". Close "My Device".

Installation Instructions

1. Insert the USB thumb drive provided by PP Systems into the USB port on the CIRAS-3 main console (we recommend USB slot 2).
2. Open "My Device" and copy the "Ciras3SCDC_Console_vxxx.cab" file from the "\USB Storage\CIRAS3SCDC_Console_SW" folder to the root directory of the console file system. Note the exact location is not important as the cab file will be erased when installation is complete.
3. Remove the USB thumb drive from console.
4. Right click on cab file and ensure it is not Write protected.
5. Double click on the cab file to execute the install. Accept all defaults by using the "OK" button.
6. When installation is complete, power off the console for a few seconds and then power it on again. The new version of the console application should start automatically.

110 Haverhill Road
Suite 301
Amesbury, MA 01913
USA

Tel: +1 978.834.0505
Fax: +1 978.834.0545

Email:
sales@ppsystems.com

URL:
www.ppsystems.com